



Ref. Certif. No.

DE 3 - 30579

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEMESYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC**CB TEST CERTIFICATE
CERTIFICAT D'ESSAI OC**

Product	Electronic measuring equipment
Name and address of the applicant	E. ZOLLER GmbH & Co. KG Gottlieb-Daimler-Str. 19 74385 Pleidelsheim, GERMANY
Name and address of the manufacturer	E. ZOLLER GmbH & Co. KG Gottlieb-Daimler-Str. 19, 74385 Pleidelsheim, GERMANY
Name and address of the factory	E. ZOLLER GmbH & Co. KG Gottlieb-Daimler-Str. 19, 74385 Pleidelsheim, GERMANY
Ratings and principal characteristics	Rated voltage: 100-120 / 200-240 V Rated frequency: 50/60 Hz Rated power: 600 VA Protection class: I Air pressure: 6-8 bar Environmental conditions: 5-35 °C, 5-90 % r.H.
Customer's Testing Facility (CTF) Stage used	CTF Stage 1
Model/type Ref.	SMILEyX-z, SMILECPT-z, SMILEyT-z y: can be 320, 420 or 620 and specified the height of the z- axis z: is used as serial number and can be 00050 to 99999
Additional information (if necessary)	Certificate DE 3 - 30417 issued 2017-04-11 is replaced by this version due to technical changes.
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010
as shown in the Test Report Ref. No. which forms part of this certificate	028-713107779-000

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

Date, 2017-11-24
CB 17 11 95184 014

Ralph Fischer



TÜV SÜD Product Service GmbH · Certification Body · Ridlerstrasse 65 · D-80339 München

Product Service